



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/559,791	01/22/2007	Jan Eriksson	3682-60	7528
23117 7590 11/14/2008 NIXON & VANDERHYE, PC 901 NORTH GLEBE ROAD, 11TH FLOOR ARLINGTON, VA 22203				
EXAMINER				
LEE, GILBERT Y				
ART UNIT		PAPER NUMBER		
3673				
MAIL DATE		DELIVERY MODE		
11/14/2008		PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/559,791

Applicant(s)

ERIKSSON, JAN

Examiner

GILBERT Y. LEE

Art Unit

3673

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-10 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-5 and 8-10 is/are rejected.
- 7) ☒ Claim(s) 6 and 7 is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 08 December 2005 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. ____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO-8508)
- Paper No(s)/Mail Date ____.

- 4) ☐ Interview Summary (PTO-413)
- Paper No(s)/Mail Date ____.
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: ____.

DETAILED ACTION

Claim Objections

1. Claim 4 is objected to because of the following informalities: "such" in line 2 should be removed. Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 1, 3-5, and 8-10 are rejected under 35 U.S.C. 102(b) as being anticipated by Bucalo (US Patent No. 4,783,109).

Regarding claim 1, the Bucalo reference discloses a mobile stretcher comprising a bed face (e.g. top face of 50) on a frame (48) provided with a head-end (e.g. right portion of 50 in Fig. 2) and a foot-end (e.g. left portion of 50 in Fig. 2), the foot-end being provided with a fixedly arranged rack (Fig. 4), which carries at least one medical instrument (e.g. 40,42) as well as that the rack is formed so that the medical instrument carried by the rack is located midwise above the foot-end (Fig. 1). Note that the Bucalo reference is capable of receiving a spineboard.

Regarding claim 3, the Bucalo reference discloses the rack being provided with two rear support legs (64, 66) fixedly arranged on each side of the frame and

substantially leaning against the bed face of the frame for carrying the instrument (Figs. 1-3).

Regarding claim 4, the Bucalo reference discloses each support leg being provided with a support part (e.g. 80 and 82 respectively) that is substantially parallel to the bed face of the frame and that extends forwards towards the head-end of the stretcher (Figs. 1-4).

Regarding claim 5, the Bucalo reference discloses a front end of at least one support part being provided with a front support leg (68), which is arranged to supportingly abut leaningly towards the bed face of the frame (Figs. 1-3).

Regarding claim 8, the Bucalo reference discloses the rack being provided with at least one transverse support member (90, 92) for supporting in parallel support parts and against which the instrument is arranged to be supported and also fixed (Figs. 1-3).

Regarding claim 9, the Bucalo reference discloses the rack being provided with at least one crash bow (e.g. 102,202) which surrounds the sides and the top side of the medical instrument (Figs. 2 and 8).

Regarding claim 10, the Bucalo reference discloses the crash bow being provided with a longitudinal protective bow (e.g. 104).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the

invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claim 2 is rejected under 35 U.S.C. 103(a) as being unpatentable over Bucalo.

Regarding claim 2, the Bucalo reference discloses the invention substantially as claimed in claim 1, including varying the height of the instrument panel to allow for passage into an ambulance entrance (Col. 6, Lines 45-54).

However, the Bucalo reference fails to disclose the height of the bottom side of said medical instrument being between 30cm and 50cm.

Discovering an optimum range of a result effective variable involves only routine skill in the art. Since applicant has not shown some unexpected result the inclusion of this limitation is considered to be a matter of mechanical expedience. It would have been obvious to one having ordinary skill in the art at the time the invention was made to make the height of the bottom side of said medical instrument be between 30cm and 50cm in the Bucalo as a matter of mechanical expedience and in order to insure passage into an ambulance entrance (Col. 6, Lines 45-54).

Allowable Subject Matter

4. Claims 6 and 7 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

The following is a statement of reasons for the indication of allowable subject matter: The closest prior art, Bucalo (US Patent No. 4,783,109) discloses the invention substantially as claimed in claims 1, 3-5, and 8-10; however, the Bucalo reference fails

to explicitly disclose the front support leg being turnably mounted. Although Hoogendoorn (US Patent No. 5,327,600) discloses turning a leg into a handle, there would be no motivation to modify fixed leg 68 of the Bucalo reference since the Bucalo reference teaches to have a "patient-loading side". There is no motivation, absent the applicant's own disclosure, to modify the Bucalo reference in the manner required by the claims.

Conclusion

5. Any inquiry concerning this communication or earlier communications from the examiner should be directed to GILBERT Y. LEE whose telephone number is (571)272-5894. The examiner can normally be reached on 8:00 - 4:30, M-F.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Patricia L. Engle can be reached on (571)272-6660. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Patricia L Engle/
Supervisory Patent Examiner, Art Unit 3673

/G. Y. L./
Examiner, Art Unit 3673